

PTR Operating Conditions

Bay Area Air Quality District Odor Allocation Project

May 12-20, 2021 Test Program

The Proton Transfer Reaction Time of Flight Mass Spectrometer (Ionicon Model 6000/X3) was operated each day using the same set of operating conditions (Setting), denoted as “Odor” in this program. Screen shots presented here denote the initial day and final day of testing. As can be seen here, the PTR instrument operating parameters and conditions are very similar each day, which demonstrates overall consistency of day to day operations and determinations of ambient sample concentrations.

May 12, 2021 Initial Operating Parameters

The screenshot shows the software interface for the Ionicon Model 6000/X3 PTR mass spectrometer. The top menu bar includes icons for file operations (New, Open, Save, Print, Exit) and a help icon. Below the menu is a toolbar with icons for setting, primary ion, transmission, and other controls. The main window displays various operating parameters in a grid format:

	Man/Ctrl	Ctrl		
PC	353.3	353.34 mbar		
p Drift	2.30	2.29 mbar		
TofLens		4.81E-5 mbar		
TOF		5.85E-7 mbar		
E/N		120 Td		
Temps	80.10 °C	79.90 °C		
SrcValve	50.0			
H ₂ O	6.0	6.00 sccm		
O ₂	0.0	0.00 sccm		
NO	0.0	0.00 sccm		
I _{hc}	4	4.0 mA		
	On/Off	On		
FCinlet	60.0	59.94 sccm		
U	FU	°C	D \rightarrow	D \leftarrow
	Us	150		145.0 V
	Uso	80		78.6 V
	Udrift	525		526.1 V

Ion Production Settings

TPS *Changed*				
		MCP	TOF	
Lens 1	12.0	13.0 V		All on <input checked="" type="checkbox"/>
Lens 2	30.0	30.0 V		Lenses <input checked="" type="checkbox"/>
Lens 3	20.0	21.0 V		
Lens 4	76.0	76.0 V		
Lens 5	70.0	70.0 V		
Lens 6	60.0	60.0 V		
Lens 7	17.0	18.0 V		
Push L	16.5	16.0 V	<input checked="" type="checkbox"/>	3 mA
Push H	790.0	790.0 V	<input checked="" type="checkbox"/>	3 mA
Pull L	86.0	86.0 V	<input checked="" type="checkbox"/>	3 mA
Pull H	700.0	700.0 V	<input checked="" type="checkbox"/>	3 mA
Grid	2400.0	2283.0 V	<input checked="" type="checkbox"/>	1 μ A
Cage	5020.0	4768 V	<input checked="" type="checkbox"/>	100 μ A
Refl. Grid	665.0	632.0 V	<input checked="" type="checkbox"/>	76 μ A
Refl. Back	900.0	855.0 V	<input checked="" type="checkbox"/>	167 μ A
MCP F	5400	5134 V	<input checked="" type="checkbox"/>	17 μ A
MCP B	2550	2536 V	<input checked="" type="checkbox"/>	247 μ A

TOF Lens Voltages and Settings

Acquisition **ACQ active**

Single Spec Time (ms)	1000	<input type="button" value="▲"/>	<input type="button" value="▼"/>
Extraction time (μs)	5.0	<input type="button" value="▲"/>	<input type="button" value="▼"/>
max Flighttime(μs)	32.0	<input type="button" value="▲"/>	<input type="button" value="▼"/>
	372.8 amu		31.25 kHz

Data Save Settings

Spec Trace Raw

Time Duration

02:00:00 Single File Duration

12 Number of Files To Store

C:\Ionicon\data

Add File Count Extension
 New ACQ for new file


<year>_<month>_<day>\
Data_<hour>_<minute>_<second>

Acquisition Settings


Mass Axis Calibration **AutoCAL done**

				<input checked="" type="checkbox"/> Cal	<input type="button" value="30 sec"/>	<input type="button" value="▲"/>	<input type="button" value="▼"/>
Mass	TimeBin						
21.0220	15965	<input type="button" value="Delete"/>	<input type="button" value="^"/>	a	15007		
203.9400	161478	<input type="button" value="Delete"/>	<input type="button" value="▼"/>	b	-52835.8		
330.8500	220131	<input type="button" value="Delete"/>	<input type="button" value="▼"/>				

Mass Axis Calibration



Hex Settings




Hydronium Ion Isotope (21.022 amu) Stability Check

May 20, 2021 PTR Settings

Final Operating Parameters

Setting	Odor			
Primary Ion	H ₃ O ⁺			
Transmission	DC			
	Man/Ctrl	Ctrl		
PC	343.5	343.52 mbar		
p Drift	2.30	2.30 mbar		
TofLens		5.35E-5 mbar		
TOF		7.91E-7 mbar		
E/N		120 Td		
Temps	80.00 °C	79.90 °C		
SrcValve	50.0			
H ₂ O	6.0	6.00 sccm		
O ₂	0.0	0.00 sccm		
NO	0.0	0.00 sccm		
Ihc	4	4.0 mA		
	On/Off	On		
FCinlet	60.0	60.01 sccm		
U	FU	°C	D \leftrightarrow	
Us	150	145.0 V		
Uso	80	78.6 V		
Udrift	525	526.1 V		

Ion Production Settings



The image shows a software interface for managing beamline optics. At the top left are icons for file operations (New, Open, Save, Print). In the center top are buttons for 'MCP' and 'TOP'. On the right side are two checkboxes: 'All on' and 'Lenses'.

Lens 1	12.0	13.0 V	All on <input checked="" type="checkbox"/>
Lens 2	30.0	30.0 V	Lenses <input checked="" type="checkbox"/>
Lens 3	20.0	21.0 V	
Lens 4	76.0	76.0 V	
Lens 5	70.0	69.0 V	
Lens 6	60.0	60.0 V	
Lens 7	17.0	18.0 V	
Push L	16.5	16.0 V	<input checked="" type="checkbox"/> 3 mA
Push H	790.0	790.0 V	<input checked="" type="checkbox"/> 2 mA
Pull L	86.0	86.0 V	<input checked="" type="checkbox"/> 3 mA
Pull H	700.0	700.0 V	<input checked="" type="checkbox"/> 3 mA
Grid	2400.0	2283.0 V	<input checked="" type="checkbox"/> 1 μ A
Cage	5020.0	4768 V	<input checked="" type="checkbox"/> 99 μ A
Refl. Grid	665.0	632.0 V	<input checked="" type="checkbox"/> 75 μ A
Refl. Back	900.0	855.0 V	<input checked="" type="checkbox"/> 167 μ A
MCP F	5400	5134 V	<input checked="" type="checkbox"/> 17 μ A
MCP B	2570	2479 V	<input checked="" type="checkbox"/> 235 μ A

TOF Lens Voltages and Settings

Acquisition **ACQ active**

Single Spec Time (ms) 1000
Extraction time (μs) 5.0 372.7 amu
max Flighttime(μs) 32.0 31.25 kHz

Data Save Settings

Spec Trace Raw

Time Duration
02:00:00 Single File Duration

12 Number of Files To Store

C:\Ionicon\data

Add File Count Extension
 New ACQ for new file

<year>_<month>_<day>\
Data_<hour>_<minute>_<second>


Acquisition Settings

Mass Axis Calibration


Cal 30 sec

Mass	TimeBin	a	b		
21.0220	15978	<input type="button" value="Delete"/>	<input type="button" value="Up"/>	15008.2	
203.9400	161502	<input type="button" value="Delete"/>	<input type="button" value="Gray"/>	b	-52828.7
330.8500	220160	<input type="button" value="Delete"/>	<input type="button" value="Down"/>		

Mass Axis Calibration Values



Hex Settings



Hydronium Isotope (21.022 amu) Stability Check